



DEPARTMENT OF HEALTH & HUMAN SERVICES

Centers for Disease Control  
and Prevention (CDC)

NIOSH Reference: TN-24876  
Mfr. Reference: NOR2006

National Institute for Occupational  
Safety and Health (NIOSH)  
National Personal Protective  
Technology Laboratory (NPPTL)  
626 Cochran Mill Road  
Pittsburgh, PA 15236-0070  
Phone: 412-386-4000  
Fax: 412-386-4051  
July 19, 2021

Edna deMedeiros  
Regulatory Compliance and Test Manager  
Honeywell International Inc.  
10 Thurber Blvd.  
Smithfield, Rhode Island 02917

Dear Edna deMedeiros:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted April 27, 2021. This request was for approval of the H801V Plus air-purifying filtering facepiece respirator with exhalation valve for protections against particulates at a N95 filter efficiency level. The complete respirator configuration is detailed on assembly matrix, file name *H801PlusSeriesAMa.xls*, revision A, dated: June 3, 2021.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired to be produced in languages other than English. Approval number TC-84A-9361 has been assigned. This respirator is approved for protection against particulates at N95 filter efficiency level.

The approval label is included as an attachment to this letter. The abbreviated label has been accepted as submitted. The cautions and limitations are listed on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

Jeffrey Peterson  
Chief, Conformity Verification and  
Standards Development Branch

Enclosures