

PHILIPS



Certificate of Conformity

AEDs associated with the HeartStart product families listed below have passed all required manufacturing test protocols and are certified to conform to their functional design. The product(s) have passed all inprocess manufacturing tests and meet all final product release specifications.

HeartStart FR2+ Defibrillator

Refurbished HeartStart FR2+ Defibrillator

HeartStart HS1 Defibrillator

HeartStart FRx Defibrillator

HeartStart FR3 Defibrillator

Required Manufacturing Test Series	Test Result
Unconfigured Mechanical Test	Pass
Unconfigured Final Test	Pass
Configured Language Test	Pass
Configured Functional Button Test	Pass
Device Battery Insertion Test	Pass
Device Simulated Use Test	Pass
Final Device Acceptance Test	Pass
Final Label Inspection	Pass

A handwritten signature in black ink that reads "Artha Nafie".

Artha Nafie
Senior Manager, Quality
Philips AED
22100 Bothell Everett Hwy, MS292
Location: (A2,A8)
Bothell, WA 98021