

吸氣抵抗測量表

(Breathing Resistance Test)

Item No.		吸氣抵抗值(kpa)	
		(Breathing resistance, kpa)	
		於 30 L/min 之空氣流量下 (At 30 liter/min)	於 95 L/min 之空氣流量下 (At 95 liter/min)
1	RC203	0.09	0.40~0.43
2	"	0.10	0.46~0.48
3	"	0.10	0.46~0.49
4	"	0.10	0.47~0.50
5	"	0.10	0.45~0.48
NOTE:			

STANDARD OF CE CERTIFICATE

Maximum inhalation permitted resistance (kpa)					
Filter Toxic Gas			Filter Fine Dust		
Class	30L/min	95L/min	Class	30L/min	95L/min
1	0.10	0.40	P1	0.06	0.21
2	0.14	0.56	P2	0.07	0.24
3	0.16	0.64	P3	0.12	0.42

OPERATING MANAGER: 

DATE: January 11, 2016